



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED

By Carol Day at 8:30 am, Oct 21, 2013

T #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).

Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.

Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 950064	NAME OF AGENCY Sedalia Police Department	DATE OF INSPECTION 10/19/2013
LOCATION OF INSTRUMENT (STREET AND CITY) 201 W. 2nd St. Sedalia, MO 65301		TIME OF INSPECTION 8:22 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 0849
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

☒ INDICATOR LIGHTS

☒ SIMULATOR SOLUTION SUPPLIER REPCO LOT # 13002 EXP. DATE 06/19/2015

☒ SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN SD2306 EXP. DATE 10/18/2014

☒ CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- ☒ 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 .095 TEST 2 .096 TEST 3 .096

☒ PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09) 1	(.10-.14)	(.15-.19)	OVER .19
----------	---------	-------------	-----------	-----------	----------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS
(USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER

SIGNATURE A. Casey Devorss	PRINT FULL NAME AARON CASEY DEVORSS
TYPE II PERMIT NUMBER/EXPIRATION DATE 230115 06.12.2015	TELEPHONE NUMBER (660) 826-8100

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13002

EXPIRATION DATE: June 19, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013
The expiration date for this lot number is June 19, 2015 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
SEDALIA POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 950064
10/19/13

TESTING OFFICER:
DEVORSS/AARON/CASEY
OFFICER I.D.: 177
PERMIT NUMBER: 230115
EXPIRATION DATE: 06/12/15
MISCELLANEOUS DATA:
TEST!

---- SUPERVISOR MODE ----

BLANK TEST	.000	09:02
INTERNAL STANDARD	VERIFIED	09:02
EXTERNAL STANDARD	.095	09:02
BLANK TEST	.000	09:03
EXTERNAL STANDARD	.095	09:03
BLANK TEST	.000	09:04
EXTERNAL STANDARD	.095	09:04
BLANK TEST	.000	09:05

N = 3
SIM. = .1
AVG. = .0956

MTNCE

Operator Signature

A. Casey Devorss 177

2208-02

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
SEDALIA POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 950064
10/19/13
09:49

---- DIAGNOSTIC CHECK ----

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS	
SAMPLE CHAMBER:	40c
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnop
qrstuvwxyz{|}~

Diagnostic

Operator Signature

A. Casey Devorss 177

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
SERIALIA POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 958264
13/19/13

ARREST TIME: 28:30

SUBJECT NAME:

NAME

DOB: 04/23/81

SEX: M

STATE/D.L.# MO/12345678

ARRESTING OFFICER:

DEVORSS/R/C

OFFICER I.D.# 177

TESTING OFFICER:

NAME

OFFICER I.D.# 177

PERMIT NUMBER: 280115

EXPIRATION DATE: 06/12/15

MISCELLANEOUS DATA

RFI TEST

--- BREATH ANALYSIS ---

RADIO INTERFERENCE

RFI

Operator Signature

A. Casey Dumas 177



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



**PERMIT
TYPE II**

AARON C DEVORSS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE **06/12/2013**

NUMBER **230115**

EXPIRES **06/12/2015**

MO 569-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

David Voshell
Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LA9-4 (RS-10)